

ACCURATE SCIENTIFIC INSTRUMENTS



BLOOD BAG TUBE SEALER

Accurate Scientific is pleased to introduce the latest Radio Frequency Sealing System.

Catalog No.	ASTS - 10
Oscillation Frequency	40.68 MHz
Radio Frequency Out put	15 W or Less
Max sealing time	Approx 1.5 to 2.5 sec.
Power Voltage	AC 230 V
Power Frequency	50 Hz
Power Consumption	100 VA or less
Size (WxDxH)	32x30x14 cm
Weight	Approx 8 kg.

FEATURES

Accurate blood bag tube is a compact equipment to seal the blood bag pilot tube by radio frequency sealing system. It takes less than 2.5 seconds to seal per tube but it is depend upon the tube Variation. Sealed tube can be easily separated by pulling the sealing section by both sides.

NB. Specifications are subject to change without prior notice.

A/04, Anand Mangal Industrial Estate, Survey No. 70/ B, Sativali Road, Valiv Vasai -(E)
Dist- Thane- 401205, Maharashtra
Tel: 0250-3200308 / 6452554- Fax: 022 28756182
Email: accurate_scientificinstruments@hotmail.co.uk
asi_lab@yahoo.com
Website: www.asilab.net